


EuroEA Elemental Analyser



AutoRun name : Zaid (6)
 Date of Analysis : 10 Feb 2019
 Time of Analysis : 12:19:05
 Analysed By : EVR
 Signed By : EVR
 Operator Group : GRP1
 Configuration : CHNS
 Calibration Type : Linear

Calibration Curve for Hydrogen



Slope	Intercept	Correlation
9.139758E-08	-0.005	0.9997628

#	Name	Type	Area	H %	KFactor	Weight (mg)
3	Acetanilide	Std	498,374	6.749	-	0.603
4	Acetanilide	Std	573,769	6.693	-	0.711
5	Acetanilide	Std	600,383	6.669	-	0.750
6	Acetanilide	Std	761,809	6.754	-	0.959
7	Acetanilide	Std	870,801	6.697	-	1.116